

Test Report

No. SH10009758/TX

Date: Jan 27, 2010

Page 1 of 5

ELIM MEDICAL TECHNOLOGY (SUZHOU) CO., LTD
409-410, MULTI-FUNCTIONAL BUILDING, HIGH-TECH VENTURE PARK, WUJIANG, JIANGSU

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : One sample of non-woven cutting in blue for disposable hospital curtain
Manufacturer : ELIM MEDICAL TECHNOLOGY (SUZHOU) CO., LTD.
Test Performed : Selected test(s) as requested by applicant
Sample Receiving Date : JAN. 16, 2010
Testing Period : JAN. 18, 2010 –JAN. 27, 2010
Test Result(s) : For further details, please refer to the following page(s).

Signed for and on behalf of
SGS-CSTC Ltd.

Abe Bian

Abe Bian (Account Executive)

COPY

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/terms_and_conditions.htm and for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, jurisdiction and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its issue. Invention only, within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated, the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



3rd Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763
t (86-21) 61402666 f (86-21) 64958763

SHTX 0081438

www.cn.sgs.com
e sgs.china@sgs.com

Test Result

I. Sample details

Materials / Color	Fabric / Blue		
Size of sample	Length: 200±2mm	Width: 160 ±2mm	32pcs
Cleaning procedure	BS EN ISO 10528	50 cycles	
Per-condition	T: 20± 2 °C H: 65±5 % at least 24 h		

II. Test Conducted

This test is conducted accordance with Type C of BS 5867-2 :2008 Specification for Fabrics for curtains, drapes and window blinds – Part2: Flammability requirements and test method is based on BS EN ISO 15025:2002.

III. Test Results

A) Before any cleaning procedure

The test methods is as per BS EN ISO 15025:2002, Procedure "A" (surface ignition, Flame application time 5s)	Lengthwise		Widthwise	
	1 (Face)	2 (Back)	1 (Face)	2 (Back)
Flame reached an edge (YES or NO)	NO	NO	NO	NO
Hole reached an edge (YES or NO)	NO	NO	NO	NO
Flaming debris (YES or NO)	NO	NO	NO	NO
Duration of flaming (s)	0	0	0	0
Duration of Afterglow (s)	0	0	0	0
Max. vertical extent of hole (mm)	58	56	54	59
Max. horizontal extent of hole (mm)	23	25	20	24

The test methods is as per BS EN ISO 15025:2002, Procedure "A" (surface ignition, Flame application time 15s)	Lengthwise		Widthwise	
	1 (Face)	2 (Back)	1 (Face)	2 (Back)
Flame reached an edge (YES or NO)	NO	NO	NO	NO
Hole reached an edge (YES or NO)	NO	NO	NO	NO
Flaming debris (YES or NO)	NO	NO	NO	NO
Duration of flaming (s)	0	0	0	0
Duration of Afterglow (s)	0	0	0	0
Max. vertical extent of hole (mm)	73	69	67	65
Max. horizontal extent of hole (mm)	29	28	34	26

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/terms_and_conditions.htm and for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, jurisdiction and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated, the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



SHTX 0081439

Attention: To check the authenticity of testing inspection report & certificate, pls. contact tel: (86-755)33071443 email: CN.Doccheck@sgs.com

The test methods is as per BS EN ISO 15025:2002, Procedure "A" (surface ignition, Flame application time 20s)	Lengthwise		Widthwise	
	1 (Face)	2 (Back)	1 (Face)	2 (Back)
Flame reached an edge (YES or NO)	NO	NO	NO	NO
Hole reached an edge (YES or NO)	NO	NO	NO	NO
Flaming debris (YES or NO)	NO	NO	NO	NO
Duration of flaming (s)	0	0	0	0
Duration of Afterglow (s)	0	0	0	0
Max. vertical extent of hole (mm)	83	76	75	73
Max. horizontal extent of hole (mm)	30	26	29	25

The test methods is as per BS EN ISO 15025:2002, Procedure "A" (surface ignition, Flame application time 30s)	Lengthwise		Widthwise	
	1 (Face)	2 (Back)	1 (Face)	2 (Back)
Flame reached an edge (YES or NO)	NO	NO	NO	NO
Hole reached an edge (YES or NO)	NO	NO	NO	NO
Flaming debris (YES or NO)	NO	NO	NO	NO
Duration of flaming (s)	0	0	0	0
Duration of Afterglow (s)	0	0	0	0
Max. vertical extent of hole (mm)	85	76	88	81
Max. horizontal extent of hole (mm)	32	30	28	31

B) After any cleaning procedure

The test methods is as per BS EN ISO 15025:2002, Procedure "A" (surface ignition, Flame application time 5s)	Lengthwise		Widthwise	
	1 (Face)	2 (Back)	1 (Face)	2 (Back)
Flame reached an edge (YES or NO)	NO	NO	NO	NO
Hole reached an edge (YES or NO)	NO	NO	NO	NO
Flaming debris (YES or NO)	NO	NO	NO	NO
Duration of flaming (s)	0	0	0	0
Duration of Afterglow (s)	0	0	0	0
Max. vertical extent of hole (mm)	57	59	54	58
Max. horizontal extent of hole (mm)	23	26	21	28

Attention: To check the authenticity of testing / inspection report & certificate, pls. contact tel: (86-755)89071443, email: CN_DocCheck@sgs.com

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/terms_and_conditions.htm and for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, jurisdiction and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated, the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



Attention: To check the authenticity of testing inspection report & certificate, pls. contact tel: (86-755)83071443 email: CN.Docecheck@sgs.com

The test methods is as per BS EN ISO 15025:2002, Procedure "A" (surface ignition, Flame application time 15s)	Lengthwise		Widthwise	
	1 (Face)	2 (Back)	1 (Face)	2 (Back)
Flame reached an edge (YES or NO)	NO	NO	NO	NO
Hole reached an edge (YES or NO)	NO	NO	NO	NO
Flaming debris (YES or NO)	NO	NO	NO	NO
Duration of flaming (s)	0	0	0	0
Duration of Afterglow (s)	0	0	0	0
Max. vertical extent of hole (mm)	66	63	59	65
Max. horizontal extent of hole (mm)	30	25	24	27

The test methods is as per BS EN ISO 15025:2002, Procedure "A" (surface ignition, Flame application time 20s)	Lengthwise		Widthwise	
	1 (Face)	2 (Back)	1 (Face)	2 (Back)
Flame reached an edge (YES or NO)	NO	NO	NO	NO
Hole reached an edge (YES or NO)	NO	NO	NO	NO
Flaming debris (YES or NO)	NO	NO	NO	NO
Duration of flaming (s)	0	0	0	0
Duration of Afterglow (s)	0	0	0	0
Max. vertical extent of hole (mm)	75	73	69	70
Max. horizontal extent of hole (mm)	36	34	39	30

The test methods is as per BS EN ISO 15025:2002, Procedure "A" (surface ignition, Flame application time 30s)	Lengthwise		Widthwise	
	1 (Face)	2 (Back)	1 (Face)	2 (Back)
Flame reached an edge (YES or NO)	NO	NO	NO	NO
Hole reached an edge (YES or NO)	NO	NO	NO	NO
Flaming debris (YES or NO)	NO	NO	NO	NO
Duration of flaming (s)	0	0	0	0
Duration of Afterglow (s)	0	0	0	0
Max. vertical extent of hole (mm)	83	78	86	80
Max. horizontal extent of hole (mm)	47	29	40	36

Criteria:

No part of any hole nor any part of the lowest boundary of any flame shall reach the top edge or either vertical edge of the sample specimen and there shall be no separation of any flaming debris from any specimen, or if the mean afterflame or afterglow times exceed 2.5 s, the fabric shall be deemed not to comply with the requirements for type "C" of this British Standard.

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/terms_and_conditions.htm and for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, jurisdiction and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.




Test Report

No. SH10009758/TX

Date: Jan 27, 2010

Page 5 of 5

Photo Appendix:



Conclusion: The submitted sample **meets** the requirements of Type C of BS 5867-2:2008.

End of Report

COPY

Attention: To check the authenticity of testing / inspection report & certificate, pls. contact tel: (66-755)83071443 email: CN.Detect@sgs.com

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/terms_and_conditions.htm and for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated, the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

